

magtivio


---

## MagSi-DNA Vegetal II

**Art.No.**

MDKT00160096

MDKT00160960



**Product Manual**

Version 5.1 | 09-12-2022

### Revision history

Revision	Release date	Remarks
1.0	24/04/2019	Initial release
5.0	08/03/2022	Updated sections 2.2 and 4.3, layout changes
5.1	09/12/2022	New company style

## Table of Contents

<b>1. General Information.....</b>	<b>1</b>
1.1 Intended Use.....	1
1.2 Kit specifications.....	1
1.3 Principle of operation.....	1
<b>2. Materials.....</b>	<b>2</b>
2.1 Kit Contents.....	2
2.2 Reagents, consumables and equipment to be supplied by the user.....	3
2.2.1 Reagents.....	3
2.2.2 Consumables and equipment for manual use or automated processing on liquid handling robots.....	3
2.2.3 Consumables for processing on the PurePrep 96 System or KingFisher™ Flex instrument.....	3
<b>3. Kit usage.....</b>	<b>4</b>
3.1 Storage Conditions.....	4
3.2 Preparation of reagents.....	4
3.3 Safety instructions.....	4
3.4 Considerations.....	4
3.5 Magnetic Separation systems.....	5
3.6 Shaker settings.....	5
3.7 Product use limitations.....	5
<b>4. Protocols.....</b>	<b>6</b>
4.1 Manual DNA extraction from plant samples (with drying step).....	6
4.2 Manual DNA extraction from plant samples with Wash Buffer III.....	7
4.3 Protocol for the PurePrep 96 System.....	8
4.3.1 PurePrep 96 software protocol file.....	8
4.3.2 Homogenization and lysis.....	8
4.3.3 Preparation of processing plates.....	8
4.3.4 Detailed instructions.....	9
4.4 Protocol for the KingFisher Flex™ Magnetic Particle Processor.....	10
4.4.1 KingFisher BindIt software protocol.....	10
4.4.2 Homogenization and lysis.....	10
4.4.3 Preparation of processing plates.....	10
4.4.4 Detailed instructions.....	11
<b>5. Troubleshooting.....</b>	<b>12</b>

## 1. General Information

### 1.1 Intended Use

**MagSi-DNA Vegetal II** is intended for Research Use Only (RUO). The kit is suited for qualified personnel only.

The kit was optimized to extract DNA from plant samples with the highest purity. The extraction chemistry was specifically developed for use with plant samples rich in fats and oils, especially seeds. While many plant DNA extractions require a dilution to further eliminate PCR inhibition, DNA isolated with MagSi-DNA Vegetal II is directly usable in downstream analysis.

The kit is intended for manual and automated isolation of genomic DNA from plant samples. Processing time for DNA extraction from 96 plant lysates is about 30 minutes. The kit requires no phenol/chloroform extraction or ethanol precipitation and eliminates the need for repeated centrifugation, vacuum filtration or column separation. It allows safe handling of samples, and is designed to avoid sample-to-sample cross-contaminations.

**MagSi-DNA Vegetal II** is suitable for automation on most liquid handling robots. The total processing time depends on the throughput and configuration of the instrument. The beads are easy to handle, have a high binding capacity and enable incubation without intensive mixing.

### 1.2 Kit specifications






The kit provides reagents for extraction of DNA from up to 10 plant seeds depending on the size, 20-50 mg fresh plant leaf or up to 10 mg lyophilized plant leaf. Purified DNA samples can be stored at 2-8°C. For long-term use, storage at -20°C is recommended. To maintain the high-molecular weight nature of the isolated DNA, it is recommended to avoid freeze-thaw cycles. Stability of lysed plant samples is dependent on the plant species. Lysed samples are typically stable for at least one day at RT, but it is recommended to proceed with DNA extraction immediately.

### 1.3 Principle of operation

Plant tissue is disrupted by mechanical homogenization and plant cell contents are released with Lysis Buffer VG containing chaotropic salts and detergents. Lysed samples should be cleared by centrifugation in order to remove cellular debris. By adding MagSi-VG III magnetic beads and adjusting binding conditions by addition of Binding Buffer U1, DNA binds to the magnetic beads while leaving impurities in solution. After magnetic separation and removal of the supernatant, the beads are washed three times to remove any residual contaminants and potential PCR inhibitors. A drying step makes sure all traces of ethanol are removed. Finally, purified DNA is eluted off the beads with Elution Buffer and can directly be used for downstream applications.

## 2. Materials

### 2.1 Kit Contents

		96 preps MDKT00160096	10 x 96 preps MDKT00160960
Lysis Buffer VG		50 mL	500 mL
Binding Buffer U1		40 mL	400 mL
MagSi-VG III		3 mL	30 mL
Wash Buffer I		60 mL	600 mL
Wash Buffer II		60 mL	600 mL
Elution Buffer		20 mL	200 mL
Product Manual		1	1

\*For DNA extraction without a drying step, Wash Buffer III can be ordered separately (Art.No. MD70041). Please contact magtivio customer support at [support@magtivio.com](mailto:support@magtivio.com) for further information.

## 2.2 Reagents, consumables and equipment to be supplied by the user

### 2.2.1 Reagents

- 80% ethanol
- Proteinase K (20 mg/mL) (optional), 10 µL per preparation (MDRE00130020 / MDRE00130200, magtivio)
- RNase (10 mg/mL) (optional), 10 µL per preparation (MDRE00150040, magtivio)

### 2.2.2 Consumables and equipment for manual use or automated processing on liquid handling robots

Product	Manual use	Automated use
Sample containers	1.5 or 2 mL microtubes	Recommended: Riplate®SW 96, PP, 2ml, (Ritter, 43001-0020) Nunc™ 96-well Polypropylene DeepWell™ Storage Plate 2.0mL, (Thermo Scientific, Cat.No. 278752)
Magnetic separation	MM-Separator M12 + 12 P Art.No. MDMG0001	MM-Separator 96 DeepWell Art.No. MDMG0013
Final container	1.5 or 2 mL microtubes	96-well microplate
Tissue homogenization	Commercial homogenizers, e.g. Geno/Grinder or TissueLyser	
Mixing	Tube Vortexer	Microplate shaker (min. 1000 RPM)
Heating	Incubator or water bath for plant cell lysis	

### 2.2.3 Consumables for processing on the PurePrep 96 System or KingFisher™ Flex instrument

Product	Art. No.	Contents
2 ml Deepwell Plate with square wells for KingFisher™/PurePrep 96	MDPL00200060	60 pieces
200 µL square-well Elution Plate for KingFisher™/PurePrep 96	MDPL00190060	60 pieces
96 well Tip-Comb for KingFisher™/PurePrep 96	MDPL00210060	60 pieces

### 3. Kit usage

#### 3.1 Storage Conditions

All components of the MagSi-DNA Vegetal II kit can be stored at room temperature (18-25°C). When stored under the conditions mentioned, the kit is stable as indicated by the expiry date on the label.

#### 3.2 Preparation of reagents

- If there is any precipitate present in the buffers, warm the buffer to 25-37°C to dissolve the precipitate before use.
- Using absolute ethanol, dilute with appropriate water to 80%. Prepare at least 60 mL 80% ethanol per 96 samples (e.g. mix 80 mL absolute ethanol with 20 mL H<sub>2</sub>O)
- Immediately before use, resuspend MagSi-VG III by vortexing for 20 seconds.

#### 3.3 Safety instructions

Take appropriate safety measures, such as wearing a suitable lab coat, disposable gloves, and protective goggles. Follow local legal requirements for working with biological materials. More information is found in the safety data sheets (SDS), available at [www.magtivio.com](http://www.magtivio.com) under each magtivio kit and kit component.

Infectious potential of liquid waste left over after using MagSi-DNA Vegetal II was not tested. Even though contamination of waste with residual infectious material is unlikely, it cannot be excluded completely. Therefore, liquid waste should be handled as being potentially infectious, and discarded according to local safety regulations.

#### 3.4 Considerations

1. It is recommended to use young plant tissue samples and keep plants in the dark to reduce polysaccharide content. In many cases lyophilized, dried material can be processed more easily and gives higher yield. Depending on plant species and sample type (seeds / leafs), the volume of the Lysis Buffer VG can be optimized. The lysis process is most efficient when using well homogenized sample material. We recommend the use of commercial homogenizers.
2. In some cases, lysis efficiency can be improved by addition of 10 µL Proteinase K (20 mg/mL).
3. If samples contain large amounts of RNA, it is recommend to add 10 µL RNase A (10 mg/mL) to the lysis mixture before incubation.
4. Elution can be performed at room temperature. Yields may be increased if elution is performed at 60°C. In most consumables elution can be carried out in ≥50 µL. Do not use less than the minimum working volume of the container used as it is essential to completely submerge the beads in elution buffer during the elution step and in order to allow magnetic separation. For some separators and sample containers, higher or lower elution volumes may be needed to contact the whole magnetic bead pellet. To avoid cross-contamination and DNA degradation, change pipette tips after each use and use nuclease-free filter-tips.

5. Lysis efficiency and DNA yield are highly dependent on the sample type. Different volumes of Lysis Buffer VG can be used to increase DNA yields. Conditions for binding have to be adjusted by taking a volume of Binding Buffer U1 that is at least equal to the volume of lysate transferred after centrifugation.

### 3.5 Magnetic Separation systems

MagSi-DNA Vegetal II has been designed for use on the MM-Separator 96 DeepWell and MM-Separator M12 + 12 P. The MM-Separator M12 + 12 P (Art.No. MDMG0001) allows simultaneous processing of up to 12 samples in 2 mL microcentrifuge tubes. For processing in 96 deepwell plates, use the MM-Separator 96 DeepWell (Art.No. MDMG0013).

For use with other magnetic separators, please contact the customer support at [support@magtivio.com](mailto:support@magtivio.com).

MagSi-DNA Vegetal II is compatible with the PurePrep 96 System and the KingFisher™ Flex Magnetic Particle Processor by Thermo Scientific™. Information of use on these instruments is described in sections 4.3 and 4.4. Software protocol files are available on request.

### 3.6 Shaker settings

The speed settings for the microplate shaker described in the protocols that follow were defined with a specific instrument and microplate. When first using a plate shaker for incubation steps, the speed settings have to be set carefully for each specific plate to prevent cross contamination and spillage. Setting the speed of the shaker can be done by loading a microplate with a volume of dyed water equal to the working volume during each step, and step-wise increasing the shaker speed until droplets are observed on the surface of the plate. Set the shaker speed lower again.

### 3.7 Product use limitations

MagSi-DNA Vegetal II is intended for research use only. Do not use for other purposes than intended. The kit components can be used only once. Do not combine components of different kits unless the lot numbers are identical. Avoid leaving bottles open to prevent contamination or evaporation of the kit reagents. Process only as many plant samples in parallel as the magnetic separator allows.



## 4. Protocols

### 4.1 Manual DNA extraction from plant samples (with drying step)

#### Before starting:

- Prepare 80% ethanol (see section 3.2 )
  - Immediately before use, resuspend MagSi-VG III by vortexing for 20 seconds
1. Homogenize up to 50 mg fresh plant sample (or <10 mg lyophilized plant sample) by mechanical disruption.
  2. Add **500 µL Lysis Buffer VG** and incubate the samples at 65°C for 30 min.  
Optional: if samples contain large amounts of RNA or if samples need to be RNA-free, we recommend to add 10 µL RNase A (~10 mg/mL) to the lysis mixture.
  3. Centrifuge for 15 min (>6.000 x g) to pellet contaminants and cell debris. Transfer 400 µL cleared lysate to a deepwell microplate or microtube.
  4. Add **400 µL Binding Buffer U1** and **30 µL MagSi-VG III**. Incubate on a shaker for 5 min at 1000 RPM.
  5. Place the samples on the magnetic separator and wait for 1 min to collect the beads. Remove supernatants.
  6. Remove the sample plate from the magnetic separator and add **600 µL Wash Buffer I** to the tubes. Incubate on a shaker for 1 min at 1000 RPM. Place the tubes in a magnetic separator and wait for 1 min to collect the beads. Remove the supernatants.
  7. Repeat step 6 one more time with **600 µL Wash Buffer II** and one time with **600 µL 80% ethanol**.
  8. Dry the beads on air for 10 min to evaporate the ethanol completely.
  9. Remove the sample plate from the magnetic separator and add **50-200 µL Elution Buffer**. Incubate on a shaker for 5 min at 1000 RPM.
  10. Place the samples on the magnetic separator and wait for 1 minute to collect the beads. Transfer the eluates to new tubes. The DNA in the eluate is now ready to use.
    - If the transferred eluates contain magnetic particles, place the tubes on the magnetic separator again, separate for 1 minute and transfer the eluates to new tubes.
    - The DNA can be eluted with different volumes of Elution Buffer (depending on the required volume for subsequent analysis).

## 4.2 Manual DNA extraction from plant samples with Wash Buffer III

\*Wash Buffer III (Art.No. MD70041) needs to be ordered separately.

### Before starting:


- Prepare 80% ethanol (see section 3.2 )
  - Immediately before use, resuspend MagSi-VG III by vortexing for 20 seconds
1. Homogenize up to 50 mg fresh plant sample (or <10 mg lyophilized plant sample) by mechanical disruption.
  2. Add **500 µL Lysis Buffer VG** ● and incubate the samples at 65°C for 30 min.  
Optional: if samples contain large amounts of RNA or if samples need to be RNA-free, we recommend to add 10 µL RNase A (~10 mg/mL) to the lysis mixture.
  3. Centrifuge for 15 min (>6.000 x g) to pellet contaminants and cell debris. Transfer 400 µL cleared lysate to a deepwell microplate or microtube.
  4. Add **400 µL Binding Buffer U1** ● and **30 µL MagSi-VG III**. Incubate on a shaker for 5 min at 1000 RPM.
  5. Place the samples on the magnetic separator and wait for 1 min to collect the beads. Remove supernatants.
  6. Remove the sample plate from the magnetic separator and add **600 µL Wash Buffer I** ● to the tubes. Incubate on a shaker for 1 min at 1000 RPM. Place the tubes in a magnetic separator and wait for 1 min to collect the beads. Remove the supernatants.
  7. Repeat step 6 one more time with **600 µL Wash Buffer II** ● and one time with **600 µL 80% ethanol**.
  8. With the samples on the magnet, slowly add **800 µL Wash Buffer III** ●. Wait for 30 seconds and carefully remove the supernatant again. Do not resuspend beads and do not exceed 60 seconds as this may cause early DNA elution. When using the kit manually, it is recommended to not treat samples with Wash Buffer III simultaneously.
  9. Remove the sample plate from the magnetic separator and add **50-200 µL Elution Buffer** ●. Incubate on a shaker for 5 min at 1000 RPM.
  10. Place the samples on the magnetic separator and wait for 1 minute to collect the beads. Transfer the eluates to new tubes. The DNA in the eluate is now ready to use.
    - If the transferred eluates contain magnetic particles, place the tubes on the magnetic separator again, separate for 1 minute and transfer the eluates to new tubes.
    - The DNA can be eluted with different volumes of Elution Buffer (depending on the required volume for subsequent analysis).

## 4.3 Protocol for the PurePrep 96 System

### 4.3.1 PurePrep 96 software protocol file





Please contact magtivio for the most recent software method files. We provide the corresponding files for direct upload on the PurePrep 96 System. Refer to the PurePrep 96 user manual regarding the upload procedure of the supplied software files to the instrument.

### 4.3.2 Homogenization and lysis

1. Homogenize up to 50 mg fresh plant sample (or <10 mg lyophilized plant sample) by mechanical disruption.
2. Add **500 µL Lysis Buffer VG**  and incubate the samples at 65°C for 30 min.  
Optional: if samples contain large amounts of RNA or if samples need to be RNA-free, we recommend to add 10 µL RNase A (~10 mg/mL) to the lysis mixture.
3. Centrifuge for 15 min (>6.000 x g) to pellet contaminants and cell debris. Transfer 400 µL cleared lysate to a 2 ml Deepwell Plate with square wells for KingFisher™/PurePrep 96.

### 4.3.3 Preparation of processing plates

#### Plate filling instructions:

Plate name	Plate type	Reagent (Kit component)	Volume	Instrument Position ("Plate")
Tip plate	2 ml Deepwell Plate with square wells for KingFisher™/PurePrep 96	Empty, for loading Tip-Comb only	N/A	1
Sample Plate	2 ml Deepwell Plate with square wells for KingFisher™/PurePrep 96	Plant lysate Binding Buffer U1  MagSi-VG III	400 µL 400 µL 30 µL	2
Wash Plate 1	2 ml Deepwell Plate with square wells for KingFisher™/PurePrep 96	Wash Buffer I 	600 µL	3
Wash Plate 2	2 ml Deepwell Plate with square wells for KingFisher™/PurePrep 96	Wash Buffer II 	600 µL	4
Wash Plate 3	2 ml Deepwell Plate with square wells for KingFisher™/PurePrep 96	80% ethanol	600 µL	5
Elution Plate	200 µL square-well Elution Plate for KingFisher™/PurePrep 96	Elution Buffer 	150 µL	8

Suitable plates can be purchased at magtivio (see section 2.2 ). We strongly recommend using only the plates which are intended to use on the PurePrep 96 System. Using unsuitable plates may result in extraction failure or instrument damage.

## 4.3.4 Detailed instructions

Follow exactly the instructions as given below. Label all plates thoroughly and unambiguously to avoid any misloading during the instrument loading procedure.

1. Prepare the “Sample Plate” for the binding step with **MagSi-VG III** and **Binding Buffer U1**. To each well of the Sample Plate already containing **400 µL plant lysate**, dispense **30 µL MagSi-VG III magnetic beads** and **400 µL Binding Buffer U1** ●.
2. Prepare “Wash Plate 1” with **Wash Buffer I**. Add **600 µL Wash Buffer I** ● to each well of the corresponding deep-well plate.
3. Prepare “Wash Plate 2” with **Wash Buffer II**. Add **600 µL Wash Buffer II** ● to each well of the corresponding deep-well plate.
4. Prepare “Wash Plate 3” with **Ethanol 80%**. Add **600 µL Ethanol 80%** to each well of the corresponding deep-well plate.
5. Prepare “Elution Plate” with **Elution Buffer**. Add **150 µL Elution Buffer** ● to each well of the corresponding square-well elution plate.
6. Switch on the PurePrep 96 System and select the protocol from the user defined protocols
7. Load all plates to the PurePrep 96 instrument on indicated positions, see section 4.3.3 (right-most column). Use the clockwise / counter clockwise buttons on the instrument to rotate the turntable to the indicated positions.

Make sure that the plates are loaded in the correct orientation (especially when using partially filled plates). Place the A1 well of each plate to the A1 mark on the instruments turntable. Make sure that the plates are fixed to the positions by the clamps.

8. Press on the Tab “Run Prog.”, select the shortcut icon for the protocol and press Run to start the protocol
9. At the end of the run remove all plates from the instrument

## 4.4 Protocol for the KingFisher Flex™ Magnetic Particle Processor

### 4.4.1 KingFisher BindIt software protocol

Please contact magtivio for the most recent BindIt software method files. We provide the corresponding files for direct upload on the KingFisher magnetic particle processors. A PDF description of the method file is included. Refer to the BindIt software manual regarding the upload procedure of the supplied software files to the instrument.

### 4.4.2 Homogenization and lysis

1. Homogenize up to 50 mg fresh plant sample (or <10 mg lyophilized plant sample) by mechanical disruption.
2. Add **500 µL Lysis Buffer VG** and incubate the samples at 65°C for 30 min.  
Optional: if samples contain large amounts of RNA or if samples need to be RNA-free, we recommend to add 10 µL RNase A (~10 mg/mL) to the lysis mixture.
3. Centrifuge for 15 min (>6.000 x g) to pellet contaminants and cell debris. Transfer 400 µL cleared lysate to a 2 ml Deepwell Plate with square wells for KingFisher™/PurePrep 96.

### 4.4.3 Preparation of processing plates

#### Plate filling instructions

Plate	Plate type	Reagent (Kit component)	Volume
Sample Plate	2 ml Deepwell Plate with square wells for KingFisher™/PurePrep 96	Plant lysate Binding Buffer U1 MagSi-VG III	400 µL 400 µL 30 µL
Wash Plate 1	2 ml Deepwell Plate with square wells for KingFisher™/PurePrep 96	Wash Buffer I	600 µL
Wash Plate 2	2 ml Deepwell Plate with square wells for KingFisher™/PurePrep 96	Wash Buffer II	600 µL
Wash Plate 3	2 ml Deepwell Plate with square wells for KingFisher™/PurePrep 96	80% ethanol	600 µL
Elution Plate	200 µL square-well Elution Plate for KingFisher™/PurePrep 96	Elution Buffer	150 µL
Tip plate	2 ml Deepwell Plate with square wells for KingFisher™/PurePrep 96	Empty, for loading Tip-Comb only	N/A

Suitable plates can be purchased at magtivio (see section 2.2 ). We strongly recommend using only the plates which are intended to use on the KingFisher Flex™ System. Using unsuitable plates may result in extraction failure or instrument damage.

#### 4.4.4 Detailed instructions

Follow exactly the instructions as given below. Label all plates thoroughly and unambiguously to avoid any misloading during the instrument loading procedure.

1. Prepare the “Sample Plate” for the binding step with **MagSi-VG III** and **Binding Buffer U1**. To each well of the Sample Plate already containing **400 µL plant lysate**, dispense **30 µL MagSi-VG III magnetic beads** and **400 µL Binding Buffer U1** ●.
2. Prepare “Wash Plate 1” with **Wash Buffer I**. Add **600 µL Wash Buffer I** ● to each well of the corresponding deep-well plate.
3. Prepare “Wash Plate 2” with **Wash Buffer II**. Add **600 µL Wash Buffer II** ● to each well of the corresponding deep-well plate.
4. Prepare “Wash Plate 3” with **Ethanol 80%**. Add **600 µL Ethanol 80%** to each well of the corresponding deep-well plate.
5. Prepare “Elution Plate” with **Elution Buffer**. Add **150 µL Elution Buffer** ● to each well of the corresponding square-well elution plate.
6. Switch on the KingFisher Flex magnetic particle processor and select the protocol from the user defined protocols
7. Start the protocol.
8. Load the plates to the instrument, following the instructions on the instrument display. Order of plates start with the tip plate and ends with the sample plate. The purification process starts immediately after loading the sample plate to the instrument.

Make sure that the plates are loaded in the correct orientation (especially when using partially filled plates). Place the A1 well of each plate to the A1 mark on the instruments turntable.

9. At the end of the method remove all plates from the instrument. Follow the instructions on the instrument display.

## 5. Troubleshooting

Problem	Possible causes	Comments and suggestions
Low DNA yield	Sample contains too low or too high amounts of plant material	- Try using larger or smaller amounts of plant material
	Incomplete lysis	- Increase incubation time for lysis - Make sure Lysis Buffer VG does not contain precipitates - Add Proteinase K (10 µL, 10 mg/mL) to the sample before incubation at 65°C
	Inefficient binding to the magnetic particles	- Use correct amounts of all reagents - Make sure the microplate shaker speed is set appropriately (see section 3.6) - Increase binding time - If samples contain large amounts of RNA, add RNase A to the lysis mixture before incubation at 65°C
	Incomplete elution	- Increase drying time for evaporation of ethanol - Increase elution time from 5 to 10 minutes - Preheat Elution Buffer to 60°C before use - Perform elution at 60°C to increase elution efficiency - Try eluting twice with 100 µL Elution Buffer
	Incomplete collection of magnetic particles	- Prolong the time-to-magnet after binding step and washing steps
Degraded or sheared DNA	Incorrect storage of the sample material	- Sample should be harvested, stored and homogenized properly - Avoid repeated thawing and freezing
Problems in downstream applications / contamination in DNA sample	Ethanol in the eluted DNA	- Increase the evaporation time for Wash Buffer II - Replace drying step with Wash Buffer III (Section 4.2 )
	Salt in the eluate (high adsorption at 230 nm)	- Make sure that wash supernatants are efficiently removed - Wash Buffers should be stored and used at RT - Repeat washing step with Wash Buffer II - Replace drying step with Wash Buffer III (Section 4.2 )
	Magnetic beads remaining in the eluate	- Place the DNA eluates in the magnetic separator again, and transfer the supernatant to a new container.

## **Magtivio B.V.**

### **Office, Lab & Production**

Daelderweg 9

6361 HK Nuth (The Netherlands)

Tel: +31-(0)45-208 4810

Fax: +31-(0)45-208 4817

E-mail: [info@magtivio.com](mailto:info@magtivio.com)

### **Ordering Information**

You can order by phone, fax, or e-mail. For a fast and efficient service please provide the following order information:

1. Delivery address
2. Invoice address
3. Purchaser contact telephone number
4. End user name and telephone number (if different)
5. Purchase order number
6. Product name and catalogue number
7. Quantity and size of products
8. Indicate if your account is VAT exempt

There is no minimum order requirement.

### **Order via e-mail:**

[order@magtivio.com](mailto:order@magtivio.com)

### **Ordering via fax or telephone:**

Tel: +31-(0)45-208 4810

Fax: +31-(0)45-208 4817